

# Declaration of Conformity

**Certificate No.:** TMC200313115-ES  
**Applicant/** **Luzhou Skinod Technology Co., Ltd**  
**Address:** Block 28, No.9 Chuangxin Road, Jianguyang District, Luzhou City  
**Manufacturer/** **Luzhou Skinod Technology Co., Ltd**  
**Address:** Block 28, No.9 Chuangxin Road, Jianguyang District, Luzhou City  
**Product Name:** **Medical infrared thermometer**  
**Trade Name:** CLOC 云和家  
**Model/Item Number :** SK-T008, SK-T008A, SK-T008B, SK-T008C, SK-T006, SK-T007, SK-T009, SK-T100, SK-T200, SK-T300  
**Rated:** Input: DC9V  
**Date and** March 14, 2020-March 17, 2020  
**Number of Test**  
**Report:** TMC200313115-E TMC200313115-S  
**EC-Directive:** Medical Devices Directive 93/42/EEC  
**Test Standard:** EN 60601-1-2:2015  
 EN 60601-1:2006+A12:2014

**Conclusion**

This Declaration of EMC Compliance has been granted to applicant based on the results of tests, performed by Laboratory of TMC Testing Services (Shenzhen) Co., Ltd. on sample of the above-mentioned product in accordance with the provisions of the relevant specific standards and the Medical Devices Directive 93/42/EEC . It is possible to use CE marking to demonstrate the compliance with this Directive.

**Place and date of issue:** Shenzhen, March 18, 2020

**TMC Testing Services (Shenzhen) Co., Ltd.**

1/F., Block A, Xinshidai Gongrong Industrial Park, No. 2,  
 Shihuan Road, Shilong Community, Shiyan Street, Baoan  
 District, Shenzhen, China  
 Tel: +86-755- 86642861  
 Email:cert@tmc-lab.com Http://www.tmc-lab.com

